

## EC Declaration of Conformity

In accordance with BS EN ISO/IEC 17050-1:2010

We: Tynetec Ltd / Aidcall (Business Units of Legrand Electric)

Of: Unit 10 Cowley Road  
Blyth Riverside Business Park  
Blyth  
Northumberland  
NE24 5TF  
England.

Declare that:

869MHz Telecare Equipment (see second page for part numbers).

Operating frequencies:

Transmit: 869.2125MHz

Receive: 869.2375MHz

In accordance with the following Directives:

2006/95/EC The Low Voltage Directive

2004/108/EC The Electromagnetic Compatibility Directive

2014/53/EU The Radio Equipment Directive

Have been designed and manufactured to the following specifications:

EN300 220-2 v3.1.1 (2016-11) / EN300 220-1 v3.1.1 (2016-11)

EN301 489-3 v1.4.1 (2002-08) / EN301 489-1 v1.9.2 (2011-09)

EN50130-4:2011 +A1:2014

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the directives.

Signed



Name

Chris Dodd

Position

Managing Director

Date

15<sup>th</sup> June 2020

List of 869MHz Telecare devices included in this Declaration:

ZXT800	Touch 2 Pendant
ZXT805	Fall Detector
ZXT810	Smoke Alarm
ZXT811	Heat Alarm
ZXT812	Carbon Monoxide Alarm
ZXT815	Flood Detector
ZXT816	Temperature Sensor
ZXT819	Hearing Impaired Smoke Alarm Kit
ZXT820	Pullcord Alarm
ZXT821	Big Button Alarm
ZXT823	Bogus Caller Alarm
ZXT830	Door Alarm
ZXT831	Property Exit Sensor
ZXT832	Door Alarm & Keyswitch
ZXT840	Sensor Controller
ZXT841	Sensor Controller with Bed Sensor
ZXT842	Sensor Controller with Chair Sensor
ZXT843	Pill Dispenser
ZXT845	Companion Epilepsy Monitor
ZXT847	Enuresis Sensor with Cotton Bed Sheet
ZXT850	Unboxed Transmitter (Pendant)
ZXT851	Unboxed Transmitter (Pullcord)
ZXT852	Unboxed Transmitter (Smoke)
ZXT861	PIR Motion Detector
ZXT865	Universal Transmitter
ZXT870	Boxed Dual Smoke/CO Transmitter